42 CFR Ch. IV (10-1-11 Edition)

transfusion and transplantation purposes must do the following:

- (1) Have available and follow written policies and protocols specifying the histocompatibility testing (that is, HLA typing, antibody screening, compatibility testing and crossmatching) to be performed for each type of cell, tissue or organ to be transfused or transplanted. The laboratory's policies must include, as applicable—
- (i) Testing protocols for cadaver donor, living, living-related, and combined organ and tissue transplants;
- (ii) Testing protocols for patients at high risk for allograft rejection; and
- (iii) The level of testing required to support clinical transplant protocols (for example, antigen or allele level).
- (2) For renal allotransplantation and combined organ and tissue transplants in which a kidney is to be transplanted, have available results of final crossmatches before the kidney is transplanted.
- (3) For nonrenal transplantation, if HLA testing and final crossmatches were not performed prospectively because of an emergency situation, the laboratory must document the circumstances, if known, under which the emergency transplant was performed, and records of the transplant must reflect any information provided to the laboratory by the patient's physician.
- (g) Documentation. The laboratory must document all control procedures performed, as specified in this section.

 $[68~{\rm FR}~3703,~{\rm Jan.}~24,~2003;~68~{\rm FR}~50724,~{\rm Aug.}~22,~2003]$

§ 493.1281 Standard: Comparison of test results.

- (a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites.
- (b) The laboratory must have a system to identify and assess patient test results that appear inconsistent with the following relevant criteria, when available:
 - (1) Patient age.
 - (2) Sex.

- (3) Diagnosis or pertinent clinical data.
- (4) Distribution of patient test results.
- (5) Relationship with other test parameters
- (c) The laboratory must document all test result comparison activities.

§ 493.1282 Standard: Corrective ac tions.

- (a) Corrective action policies and procedures must be available and followed as necessary to maintain the laboratory's operation for testing patient specimens in a manner that ensures accurate and reliable patient test results and reports.
- (b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur:
- (1) Test systems do not meet the laboratory's verified or established performance specifications, as determined in §493.1253(b), which include but are not limited to—
- (i) Equipment or methodologies that perform outside of established operating parameters or performance specifications;
- (ii) Patient test values that are outside of the laboratory's reportable range of test results for the test system: and
- (iii) When the laboratory determines that the reference intervals (normal values) for a test procedure are inappropriate for the laboratory's patient population.
- (2) Results of control or calibration materials, or both, fail to meet the laboratory's established criteria for acceptability. All patient test results obtained in the unacceptable test run and since the last acceptable test run must be evaluated to determine if patient test results have been adversely affected. The laboratory must take the corrective action necessary to ensure the reporting of accurate and reliable patient test results.
- (3) The criteria for proper storage of reagents and specimens, as specified under §493.1252(b), are not met.